



BD Influx Cell Sorter

Technical Specifications

The BD Influx™ cell sorter is a flexible flow cytometry platform that adapts to a researcher's application requirements. A modular architecture and a combination of detection capabilities, hands-on controls, and high performance allow researchers to configure the BD Influx system to various site and application needs.

The BD Influx system can handle a throughput rate of up to 200,000 events per second.

Optics

Excitation Optics

Optical Platform

Lasers are mounted on a standard optical breadboard. Laser beams are aligned independently by adjustable mirrors. Each laser beam has its own final focus lens mounted on a dedicated translational stage for fine adjustments. For systems with more than five lasers, the laser paths have one or more collinear beams through the steering optics and focusing lens, but intercept the stream at up to seven spatially separated points.

Lasers

A selection of laser wavelength and power is available. New laser combinations may require custom engineering of the optical platform.

Power Out of the Laser Head

355 nm: >100 mW, >250 mW
 405 nm: >100 mW
 445 nm: >100 mW
 457 nm: >300 mW
 488 nm: >200 mW
 515 nm: >100 mW
 532 nm: >150 mW
 561 nm: >75 mW, >150 mW
 594 nm: >100 mW
 640 nm: >120 mW
 785 nm: >40 mW

Beam-shaping optics provide elliptical beam spots for selected laser paths (3:1 ratio with a typical beam height of 15–20 μm).

Emission Optics

Emission light is collected through a 20X, 0.6 NA microscope objective (90°). Light is focused on three, five, or seven spatially separated mirror pinholes depending on the number of lasers. Modular detector blocks allow for a user-defined detection configuration (see the *BD Influx Cell Sorter Filter Guide*).

Simultaneous video observation of the stream, the pinholes, and the laser intercepts allows for fast and intuitive alignment.

Regular forward scatter is detected using 75- and 50-mm lenses, an aperture, and a photomultiplier tube. Resolution for the standard forward scatter detector is >0.5 μm (measured using beads). Collection angle is 2–17°.

Side scatter is collected through the 90° collection lens and measured using a photomultiplier tube. Side scatter resolution is >0.2 μm (measured using beads and 0.1- μm filtered sheath fluid).

Fluidics

General Operation

- Laboratory air pressure and/or vacuum can be used for operation (regulated at 90 psi, 6.2 bar).
- An optional air pressure supply and/or vacuum pump is available.
- Sheath pressure is adjustable from 1–90 psi (0.07–6.2 bar).

Fluidics Reservoirs

Autoclavable 7-L sheath and waste containers, equipped with pressure and vacuum readout, are provided.

Fluidics Control

- Sheath, sample, and boost pressure can be individually adjusted.
- A sample flow fine adjustment is provided for precise regulation of sample flow.
- Purge, pulse, rinse, and run buttons are provided for quick stream startup and bubble removal.

Replaceable Fluidics Path

- The fluidics path, including the nozzle assembly, can be exchanged. There are no inline valves. Only pinch valves are used.
- The sample line can also be exchanged.

Bubble Detector

A bubble detector in the sample line detects air bubbles from the sample tube and stops sample flow when the sample tube is empty, preventing air bubbles from reaching the nozzle assembly.

Sample Input

12 x 75-mm tubes, polypropylene

Temperature Control

Sample tubes can be cooled or heated by an optional circulating water bath.

Performance

Fluorescence Sensitivity

Measured using SPHERO™ Rainbow Calibration Particles (RCP-30-5A) according to the manufacturer's specifications:

Sheath pressure: 33 psi
 Drop drive: On, ~68 kHz
 Excitation: 488 nm, 200 mW
 Emission: 530/40 nm for FITC
 580/30 nm for PE

FITC: 125 molecules of equivalent soluble fluorochrome (MESF-FITC)
 PE: 125 molecules of equivalent soluble fluorochrome (MESF-PE)

Fluorescence Resolution

Measured using propidium iodide (PI)-stained chicken erythrocyte nuclei (CEN):

Sheath pressure: 33 psi
 Drop drive: On, ~68 kHz
 Excitation: 488 nm, 200 mW
 Emission: 610/20 nm for PI

Coefficient of variation (CV) of PI: <3%, full G_0/G_1 peak

Fluorescence Linearity

Measured using propidium iodide PI-stained CEN:

Sheath pressure: 33 psi
 Drop drive: On, ~68 kHz
 Excitation: 488 nm, 200 mW
 Emission: 610/20 nm for PI

Doublet/singlet ratio: 1.95–2.05

Sort Performance

Drop Drive Frequency

Adjustable 9–180 kHz

Purity and Yield

At 60 psi and 100 kHz with an average threshold rate of 25,000 events per second, a four-way sort achieved a purity of 98% and a yield >80% of Poisson's expected yield for all four populations. Higher threshold rates of up to 70,000 events per second can be achieved without affecting purity; however, yield will decrease based on Poisson statistics.

Viability

As shown in published literature, sorts performed using murine¹⁻⁴ and human⁵ cells and/or cell lines⁶ demonstrated good recovery and viability in several experimental systems. Optimal sort conditions need to be established for different cell types.

Nozzles

Supplied: 70, 86, 100, and 140 μm

Optional: 200 μm

Sort Collection Devices

All collection devices are designed to fit on the Computerized Cell Deposition Unit (CCDU). The CCDU is standard on all instruments.

- Two-way sorting: Microtubes; 12 x 75-mm, 15-mL, and 50-mL tubes
- Three-way sorting: One 50-mL tube and two 12 x 75-mm tubes
- Four-way sorting: Microtubes, 12 x 75-mm tubes
- Plates and slides: 6, 24, 48, 96, and 384-well plates; slides; and user-defined collection devices

Sample Collection Cooling

Water recirculator for refrigeration or heating of collection tubes (optional).

Sort Monitoring

- Live video feed of waste collection and side streams.
- Live video feed of breakoff point.
- Drop-delay determination is achieved with BD FACSTM Accudrop technology. The drop-delay value can be adjusted while viewing BD Accudrop beads which are illuminated by a red diode laser in the center and side sort streams.

Signal Processing

Data Acquisition Channels

- Up to 5-laser systems: 16 channels, usually 14 colors plus forward and side scatter
- Systems with 6 lasers or more: 24 channels, usually 22 colors plus forward and side scatter

Signal Processing

- 16-bit analog-to-digital conversion, 65,536 channels
- Parallel data stream with channel ID and integrity check
- Less than 1 correlation error per 10⁸ events

Acquisition Rate

Dead time is 0 μs . The maximum throughput rate is 200,000 events per second, independent of the number of parameters.

Fluorescence Compensation

- Up to 5-laser systems: 16 x 16 digital compensation matrix.
- Systems with 6 lasers or more: 24 x 24 digital compensation matrix.
- Compensated parameters are being added to the bus as separate parameters.

Pulse Processing

- All signals are height (peak) by default.
- Pulse processor electronics add area and width measurements for a maximum of 8 parameters to the bus.
- Width measurement on the trigger parameter is standard.

Time

Time can be correlated to any parameter for kinetic experiments or other applications.

Channel Threshold

- Any parameter can be used as the threshold from the primary laser.
- Lasers and detectors can easily be switched to change laser sequence.

Workstation

Workstation

PC workstation with Intel® Quad Core (3.06 GHz or faster), Windows® 7 Professional 64-bit operating system

Memory

6 GB RAM

Data Storage

- 2 x 500 GB hard drives, RAID 1 (mirrored) configured
- 16x CD/DVD +/- RW
- 2 USB 2.0

Networking

10/100/1000 Gigabit Ethernet

Monitor

20-inch LCD, 1280x1024 resolution

Data File Structure

Flow cytometry standard (FCS) 3.0

Software

BD FACSTM Software v1.0 or later

Installation Requirements

Dimensions (H x D x W)

Cytometer: 72 x 48 x 62 in.
(183 x 122 x 158 cm)

Cytometer with HEPA-filtered enclosure:* 86 x 50 x 77 in.
(218 x 127 x 196 cm)

Operational footprint (D x W):
84 x 84 in. (214 x 214 cm)

*A minimum of 102.4 in. (260 cm) floor-to-ceiling height is required during installation of the HEPA-filtered enclosure.

Weight

Cytometer: 229 lb (104 kg)

Cytometer with HEPA-filtered enclosure:
441 lb (200 kg)

Temperature Operating Range

59–77°F (15–25°C)

Heat Dissipation

6,200 BTU/h maximum, dependent on laser choice

Electrical Requirements

North America: 120 ± 10% VAC, 50/60 Hz, 15 A

Outside North America: 230 ± 10% VAC, 50/60 Hz, 10 A

Japan: 100 VAC, 50/60 Hz, 15 A

BD requires two dedicated circuits for the cytometer and computer system with a dedicated AC source not shared with any other equipment. Multiple peripherals may require an additional line. An appropriate transformer will be delivered depending on region-specific settings. Optional air and vacuum supply require an additional line.

Power

Power requirements are between 1,080 watts (3 lasers, no HEPA-filtered enclosure) and 1,800 (5 lasers with HEPA-filtered enclosure). Standby status is not applicable.

Humidity

55% ± 10% RH

Noise

<80 dBA from all running equipment

Air Supply

90 psi (6.2 bar) regulated. The source of compressed air must deliver clean (less than 5 ppm), dry-filtered (oil-free) air and stable pressures.

Vacuum Supply

A vacuum supply is delivered with the system. A laboratory vacuum supply between 5 and 15 in. Hg at 1 CFM can also be used.

Options

Small particle

Small particle with polarization

HEPA-filtered enclosure

Aerosol evacuation

Sample temperature control

Polarization (without small particle)

Air compressor

Regulatory Status

CE marked for electrical safety (Europe)

UL Standard for electrical safety (USA)

CSA for electrical safety (Canada)

Class I (1) laser product per CDRH regulations and EN/IEC 60825-1

References

1. Chen HL, Pistollato F, Hoepfner DJ, Ni HT, McKay RDG, Panchision DM. Oxygen tension regulates survival and fate of mouse central nervous system precursors at multiple levels. *Stem Cells*. 2007;25:2291-2301.
2. Kumar R, Fossati V, Israel M, Snoeck HW. Lin-Sca1+kit-bone marrow cells contain early lymphoid-committed precursors that are distinct from common lymphoid progenitors. *J Immunol*. 2008;181:7507-7513.
3. Mar L, Hoodless PA. Embryonic fibroblasts from mice lacking Tgfr were defective in cell cycling. *Mol Cell Biol*. 2006;26:4302-4310.
4. Seamons A, Perchellet A, Gorman J. Endogenous myelin basic protein is presented in the periphery by both dendritic cells and resting B cells with different functional consequences. *J Immunol*. 2006;177:2097-2106.
5. Pistollato F, Chen HL, Schwartz PH, Basso G, Panchision DM. Oxygen tension controls the expansion of human CNS precursors and the generation of astrocytes and oligodendrocytes. *Mol Cell Neurosci*. 2007;35:424-435.
6. Panchision DM, Chen HL, Pistollato F, Papini D, Ni HT, Hawley TS. Optimized flow cytometric analysis of central nervous system tissue reveals novel functional relationships among cells expressing CD133, CD15, and CD24. *Stem Cells*. 2007;25:1560-1570.

Class I (1) laser product.

For Research Use Only. Not for use in diagnostic or therapeutic procedures.

Windows is a registered trademark of Microsoft Corporation.

Intel is a registered trademark of Intel Corporation.

SPHERO is a trademark of Spherotech, Inc.

BD, BD Logo and all other trademarks are property of Becton, Dickinson and Company. © 2011 BD