

Protocol MSU_MSMC_007, Version 1.0

Methoxyamination and trimethylsilylation of amino acids and organic acids for GC/MS analysis

Last updated on March 28, 2019 by A. Daniel Jones

Reagents

Methoxyamine hydrochloride (Sigma-Aldrich #89803)

Pyridine (dry), in ampules (Fisher #06-719-372)

N-Methyl-*N*-trimethylsilyltrifluoroacetamide (MSTFA) containing 1% trimethylsilyl chloride (TMSCl), in sealed glass ampules (Sigma-Aldrich #69478)

Supplies

1.7-mL polypropylene microcentrifuge tubes with locking caps (VWR #490016-245)

Calibrated 1000- μ L pipetter and pipet tips

Calibrated 10- μ L pipetter and pipet tips

Vortexer

Ultrasound water bath

Labeled amber autosampler vials (BMB Stores #06718439 with low volume (250 μ L) glass inserts (BMB Stores #056005203N) and PTFE-lined screw caps (BMB Stores #06718904)

Oven or heated block

Analytical balance (to 0.1 mg)

Spatula, precleaned

Samples

Use extracts of one of the following (after evaporation of solvents to dryness, typically in a screw cap vial; can use autosampler vials with inserts; extraction details are in separate SOPs):

blood serum or plasma (30 μ L)

urine (50-100 μ L)

Cell cultures (10^7 cells)

Cell culture media (50 μ L)

Homogenized tissue (plant, animal, or other): 2-25 mg of tissue

Procedure

1. Set oven (or heated block) temperature to 60°C.
2. Prepare labels for microcentrifuge tubes and GC vials
3. Use a spatula to weigh 0.040 g of methoxyamine hydrochloride into a 1.7 mL microcentrifuge tube labeled as "40 mg/mL methoxyamine-HCl in pyridine"
4. Transfer 1000 μ L of dry pyridine into the above microcentrifuge tube; seal the tube.
5. Vortex briefly, then ultrasonicate for 15 minutes; ensure that all of the solid has dissolved before proceeding.
6. Add 100 μ L of the methoxyamine-HCl/pyridine solution to each dried sample, blank, QC sample, and calibration standard.
7. Heat tubes at 60°C for 12-24 hours; allow tubes to cool to room temperature.
8. Add 100 μ L of MSTFA + 1% TMSCl and seal the tube.
9. Heat tubes at 60°C for 12-24 hours; allow tubes to cool to room temperature.
10. Transfer 50 μ L of each reaction mixture to an autosampler vial equipped with low-volume insert; cap the vial and transfer it to the GC/MS autosampler.
11. Transfer the remaining reaction materials to a separate labeled autosampler vial (no insert necessary) for storage.